Datasheet for ABIN4883958

# Norrie Disease (Pseudoglioma) ELISA Kit

Go to Product page

## Overview

| | |
|---|---|
| Quantity: | 96 tests |
| Target: | Norrie Disease (Pseudoglioma) (NDP) |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.41-100 ng/mL |
| Minimum Detection Limit: | 0.41 ng/mL |
| Application: | ELISA |

## Product Details

| | |
|---|---|
| Purpose: | Human Norrin ELISA Kit for Serum, Plasma, and Cell Culture Supernatants. |
| Sample Type: | Plasma, Cell Culture Supernatant, Serum |
| Analytical Method: | Quantitative |
| Detection Method: | Colorimetric |
| Specificity: | This ELISA antibody pair recognizes Human Norrin. |
| Characteristics: | • Strip plates and additional reagents allow for use in multiple experiments<br>• Quantitative protein detection<br>• Establishes normal range<br>• The best products for confirmation of antibody array data |
| Components: | • Pre-Coated 96-well Strip Microplate<br>• Wash Buffer<br>• Stop Solution |

 Copyright antibodies-online. All rights reserved.

## Product Details

- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 µL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

## Target Details

| | |
|---|---|
| Target: | Norrie Disease (Pseudoglioma) (NDP) |
| Alternative Name: | Norrin (NDP Products) |
| Gene ID: | 4693 |
| UniProt: | Q00604 |
| Pathways: | Sensory Perception of Sound |

## Application Details

| | |
|---|---|
| Application Notes: | Recommended Dilution for serum and plasma samples2 fold |
| Sample Volume: | 100 µL |
| Plate: | Pre-coated |
| Protocol: | 1. Prepare all reagents, samples and standards as instructed in the manual.<br>2. Add 100 µL of standard or sample to each well.<br>3. Incubate 2.5 h at RT or O/N at 4 °C.<br>4. Add 100 µL of prepared biotin antibody to each well.<br>5. Incubate 1 h at RT.<br>6. Add 100 µL of prepared Streptavidin solution to each well.<br>7. Incubate 45 min at RT.<br>8. Add 100 µL of TMB One-Step Substrate Reagent to each well.<br>9. Incubate 30 min at RT.<br>10. Add 50 µL of Stop Solution to each well. |

## Application Details

| | |
|---|---|
| | 11. Read at 450 nm immediately. |
| Restrictions: | For Research Use only |

## Handling

| | |
|---|---|
| Storage: | -20 °C |
| Storage Comment: | The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C. |
| Expiry Date: | 6 months |

Copyright antibodies-online. All rights reserved.